

NOV 08 2010

VERIFICATION OF CALIBRATION REPORT*of DataMaster cdm Breath Test Instrument**State of Alaska**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130295</u> ✓	
Name <u>JOE CORBETT</u>	ID# <u>4971</u> ✓	Date <u>10/18/10</u> ✓	
A Agency <u>Bethel Police Department</u>		Phone # <u>545-0139</u> cell	
Instrument Location <u>BETHEL PD</u>			
B Alco S/N <u>X301563</u> ✓		Target Value <u>.079</u>	High Pressure <u>600 PSI</u>
Alco Test Values		<u>.080</u> 1 st Alco	<u>.080</u> 2 nd Alco
Signature <u>[Signature]</u>			
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130295

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

OCTOBER 18, 2018

OPERATOR'S NAME:

CORBETT/JOE

OPERATOR'S NUMBER: 4971

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.I.L. #: 1234567

DEPT/AGENCY: BET1

CASE/REPORT: 10-10-10 UOC

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: X301563

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.72 in

ALCO TARGET .078 21:41

BLANK TEST .000 21:42

INTERNAL STANDARD VERIFIED 21:42

ALCO TV 29.72 in .000 21:43

BLANK TEST .000 21:43

SUBJECT SAMPLE .000 21:44

BLANK TEST .000 21:45

ALCO TV 29.72 in .000 21:45

BLANK TEST .000 21:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130295

OCTOBER 18, 2018

TIME 21:38

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.72 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢